

Results of Proficiency Test

Biogasoline E10

May 2015

Organised by: Institute for Interlaboratory Studies
Spijkenisse, the Netherlands
Authors: ing. L. Dijkstra
Correctors: dr. R.G. Visser, ing. R.J. Starink & ing. L. Sweere
Report: iis15B03

August 2015

CONTENTS

1	INTRODUCTION	3
2	SET UP.....	3
2.1	ACCREDITATION.....	3
2.2	PROTOCOL.....	3
2.3	CONFIDENTIALY STATEMENT.....	3
2.4	SAMPLES	4
2.5	STABILITY OF THE SAMPLES	5
2.6	ANALYSES	5
3	RESULTS.....	5
3.1	STATISTICS	6
3.2	GRAPHICS	6
3.3	Z-SCORES.....	7
4	EVALUATION	7
4.1	EVALUATION PER TEST	7
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES	11
4.3	COMPARISON OF THE PROFICIENCY TEST OF MAY 2015 WITH PREVIOUS PTS	12

Appendices:

1.	Data, statistical results and graphic results.....	13
2.	z-scores distillation ASTM D86.....	42
3	Number of participants per country	43
4.	Abbreviations and literature	44

1 INTRODUCTION

Since 2009, the Institute for Interlaboratory Studies organizes a proficiency test for the analysis of Biogasoline E10, in accordance with the latest applicable version of EN228 specification. During the annual proficiency testing program 2014/2015, it was decided to continue the round robin for the analysis of Biogasoline E10.

This interlaboratory study, in total 43 laboratories in 18 different countries has participated. See appendix 3 for the number of participants per country. In this report, the results of the 2015 Biogasoline E10 proficiency test are presented and discussed. This report is also electronically available through the iis internet site www.iisnl.com.

2 SET UP

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organiser of this proficiency test. The sample analyses for fit-for-use and homogeneity testing were subcontracted. In this proficiency test, the participants received, depending on their registration, 1*1 litre Biogasoline E10 (labelled #15060) and/or 1*1 litre Biogasoline E10 (\pm 750 mL filled, labelled #15061 for DVPE only) and/or 1*1 litre Biogasoline E10 (labelled #15062 for RON/MON only)

Participants were requested to report rounded and unrounded results. The unrounded results were preferably used for statistical evaluation.

2.1 ACCREDITATION

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, is accredited in agreement with ISO/IEC 17043:2010 (R007), since January 2000, by the Dutch Accreditation Council (Raad voor Accreditatie). This PT falls under the accredited scope. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by sending out questionnaires.

2.2 PROTOCOL

The protocol followed in the organisation was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of April 2014 (iis-protocol, version 3.3). This protocol can be downloaded via the FAQ page of the iis internet site www.iisnl.com.

2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and are for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

2.4 SAMPLES

The necessary sample material of about 200 litres of Biogasoline E10 was purchased at a local petrol station. From this batch, after homogenisation, 69 amber glass bottles of 1 litre for the main sample (labelled #15060) and 39 amber glass bottles for RON/MON only (labelled #15062) were filled. From the same batch another 60 brown glass bottles of 1 litre were filled with approx. 750 mL especially for Dry Vapour Pressure Equivalent (labelled #15061).

The homogeneity of the subsamples #15060 and #15062, from the same batch, was checked by determination of Density at 15°C in accordance with ASTM D4052 on 8 stratified randomly selected samples. The homogeneity of the subsamples #15061 was checked by determination of Dry Vapour Pressure Equivalent in accordance with ASTM D5191 on 8 stratified randomly selected samples.

	Density at 15°C in kg/m ³
Sample #15060-1	737.65
Sample #15060-2	737.58
Sample #15060-3	737.67
Sample #15060-4	737.66
Sample #15062-1	737.65
Sample #15062-2	737.65
Sample #15062-3	737.57
Sample #15062-4	737.68

table 1: homogeneity test results of subsamples #15060 and #15062

	DVPE in psi
Sample #15061-1	11.2
Sample #15061-2	11.3
Sample #15061-3	11.3
Sample #15061-4	11.4
Sample #15061-5	11.4
Sample #15061-6	11.4
Sample #15061-7	11.4
Sample #15061-8	11.4

table 2: homogeneity test results of subsamples #15061

From the above test results, the repeatabilities were calculated and compared with 0.3 times the corresponding reproducibilities in agreement with the procedure of ISO 13528, Annex B2 in the next table:

	Density at 15°C in kg/m ³	DVPE in psi
r (sample #15060)	0.11	--
r (sample #15061)	--	0.16
reference test	ISO12185:96	D5191:13
0.3*R (Reference)	0.45	(0.11)
r (Reference)	(0.40)	0.21

table 3: repeatabilities of the subsamples #15060 and #15061

The calculated repeatability of the Density was less than 0.3 times the reproducibility of the corresponding reference method. Due to the strict reproducibility requirements of ASTM D5191:13, the calculated repeatability for DVPE was compared with repeatability of the corresponding reference method. The calculated repeatability of DVPE was less than the repeatability of the corresponding reference method.

Therefore, homogeneity of the subsamples #15060, #15061 and #15062 was assumed.

To the participants, depending on their registration, 1*1 litre of sample #15060 and/or 1*1 litre (± 750 mL filled) of sample #15061 and/or 1*1 litre of sample #15062 were sent on April 22, 2015.

2.5 STABILITY OF THE SAMPLES

The stability of Gasoline, packed in the brown glass bottles, was checked. The material was found sufficiently stable for the period of the proficiency test.

2.6 ANALYSIS

The participants were requested to determine on sample #15060: API gravity, Aromatics (by GC), Benzene, Copper Strip Corrosion 3hrs/50°C, Density at 15°C, Distillation, Doctor test, Existence Gum, FIA (olefins & aromatics), Lead as Pb, Manganese as Mn, Mercaptan Sulphur as S, Oxidation Stability, Olefins (by GC), Oxidation Stability, Oxygenates, Oxygen and Sulphur.

On sample #15061 the participants were requested to determine TVP and to calculate DVPE only (in accordance with ASTM D5191 and EPA requirements). The participants were requested to determine RON/MON on sample #15062.

To get comparable results a detailed report form, on which the units were prescribed as well as the required standards and a letter of instructions were prepared and made available on the data entry portal www.kpmd.co.uk/sgs-iis/. The detailed report form was also made available for download on the iis website www.iisnl.com. A SDS and a form to confirm receipt of the samples were added to the sample package.

3 RESULTS

During four weeks after sample despatch, the results of the individual laboratories were gathered. The original data are tabulated per determination in the appendix of this report. The laboratories are presented by their code numbers

Directly after the deadline, a reminder fax was sent to those laboratories that had not reported results at that moment.

Shortly after the deadline, the available results were screened for suspect data. A result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the results. Additional or corrected results are used for data analysis and original results are placed under 'Remarks' in the result tables in appendix 1.

3.1 STATISTICS

Statistical calculations were performed as described in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' (iis-protocol, April 2014 version 3.3). For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded results. Results reported as '<...' or '>...' were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either 'unknown', 'OK', 'suspect' or 'not OK'. After removal of outliers, this check was repeated. Not all data sets proved to have a normal distribution, in which cases the statistical evaluation of the results should be used with due care.

In accordance to ISO 5725 (1986 and 1994) the original results per determination were submitted subsequently to Dixon, Grubbs and Rosner outlier tests. Outliers are marked by D(0.01) for the Dixon test, by G(0.01) or DG(0.01) for the Grubbs test and by R(0.01) for the Rosner General ESD test (see appendix 4, no.16). Stragglers are marked by D(0.05) for the Dixon test, by G(0.05) or DG(0.05) for the Grubbs test and by R(0.05). Both outliers and stragglers were not included in the calculations of the averages and the standard deviations.

For each assigned value, the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirement based on the target reproducibility in accordance with ISO13528. When the uncertainty passed the evaluation, no remarks are made in the report. However, when the uncertainty failed the evaluation it is mentioned in the report and it will have consequences for the evaluation of the test results.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported analysis results are plotted. The corresponding laboratory numbers are on the X-axis. The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected standard. Outliers and other data, which were excluded from the calculations, are represented as a "x". Accepted data are represented as a triangle. Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms (see appendix 4; nos 13 and 14). Also a normal Gauss curve was projected over the Kernel Density Graph.

3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements, e.g. ASTM reproducibilities, the z-scores were calculated using a target standard deviation. This target standard deviation was calculated from the literature reproducibility by division with 2.8.

When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated according to:

$$z_{(\text{target})} = (\text{result} - \text{average of PT}) / \text{target standard deviation}$$

Absolute values for $z < 2$ are very common and absolute values for $z > 3$ are very rare. Therefore, the usual interpretation of z-scores is as follows:

$ z < 1$	good
$1 < z < 2$	satisfactory
$2 < z < 3$	questionable
$3 < z $	unsatisfactory

4 EVALUATION

In this proficiency test, no problems were encountered during the dispatch of the samples to the participants.

Three laboratories reported the test results after the final reporting date and two laboratories did not report any result at all. Not all laboratories were able to perform all analyses requested. Finally, 41 laboratories did report 713 numerical results. Observed were 20 outlying results, which is 2.8%. In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

4.1 EVALUATION PER TEST

In this section, the results are discussed per sample and per test. The specified test methods and requirements were taken into account for explaining the observed differences when possible and applicable. These methods are also in the tables together with the reported data. The abbreviations, used in these tables, are listed in appendix 3.

In the iis PT reports, ASTM methods are referred to with a number (e.g. D2086) and an added designation for the year that the method was adopted or revised (e.g. D2086-08). If applicable, a designation in parentheses is added to designate the year of reapproval (e.g. D2086-08 (2013)). In the results tables of Appendix 1 only the method number and year of adoption or revision will be used.

Not all original data sets proved to have a normal Gaussian distribution. These are referred to as “not OK” or “suspect”. The statistical evaluation of these data sets should be used with due care, see also paragraph 3.1.

For sample #15060

API gravity: This determination was not problematic. One statistical outlier was observed. However, the calculated reproducibility after rejection of the statistical outlier is in good agreement with the requirements of ASTM D4052:11.

Aromatics by GC: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in full agreement with the requirements of ISO22854:14. Six participants used method ASTM D5769. This method is not equivalent to ISO22854:14. Two ASTM D5769 test results appeared to be statistical outliers. When the ISO22854 test results were evaluated separately, a higher aromatics content is found with a smaller standard deviation than for the full data set.

Benzene: This determination was problematic. No statistical outliers were observed. However, the calculated reproducibility is not in agreement with the requirements of ISO22854:14.

Copper strip: No problems have been observed, all reporting participants agreed on a test result of 1.

Density at 15°C: This determination is not problematic. One statistical outlier was observed. However, the calculated reproducibility after rejection of the statistical outlier is in good agreement with the requirements of ISO12185:96 and of ASTM D4052:11.

Distillation: The determination of the distillation was not problematic. In total five statistical outliers were observed. When compared against the automated mode requirements of ASTM D86:12, the calculated reproducibilities for IBP, 10% rec, 50% rec, 90% rec, FBP, % evap at 100°C and % evap at 150°C after rejection of the statistical outliers, are all in agreement. When compared against the manual mode, the calculated reproducibilities for IBP, 10% rec, 50% rec, 90% rec, % evap at 70°C and % evap at 150°C after rejection of the statistical outlier, are all in agreement with requirements of ASTM D86:12.

Doctor test: No problems have been observed, all reporting participants agreed on a test result of “negative”.

Existent Gum: This determination was not problematic. Three statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with requirements of ASTM D381:12.

- FIA Olefins: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ASTM D1319:14. The large spread may be caused by not or wrongly correcting the results for the (high) oxygenate content.
- FIA Aromatics: This determination was problematic. No statistical outliers were observed. However, the calculated reproducibility is not in agreement with the requirements of ASTM D1319:14. The large spread may be caused by not or wrongly correcting the results for the (high) oxygenate content.
- Lead: The consensus value of the group was below the application range (2.5 – 25 mg/L) of ASTM D3237:12. Therefore, no significant conclusions were drawn.
- Manganese: The consensus value of the group was below the application range (0.25 – 40 mg/L) of ASTM D3831:12. Therefore, no significant conclusions were drawn.
- Mercaptans: The consensus value of the group was below the application range (0.0003 – 0.01 %M/M) of ASTM D3227:13. Therefore, no significant conclusions were drawn.
- Olefins by GC: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in good agreement with the requirements of ISO22854:14.
- Oxidation Stability: All participants, except one, agreed that the Oxidation Stability is >360 minutes.
- Ethanol: This determination may be problematic depending on the test method used. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ISO22854:14 or EN1601:14 or EN13132:00, but it is in agreement with the requirements of ASTM D5599:10. When the reproducibility of the used test method differs significantly from the reproducibility of the reference test, the laboratory is advised to recalculate its z-score as described in paragraph 3.3.
- Ethers: The consensus value of the group was near or below the application range of ISO22854:14 (min. 0.2 %V/V). Therefore, no significant conclusions were drawn.

Other oxygenates: The consensus values for the various oxygenates were near or below the application range of ISO22854:14 (min. 0.2 %V/V). Therefore, no significant conclusions were drawn.

Oxygen: This determination was problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the requirements mentioned in EN228:13 (note k).

Sulphur: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ISO20846:11.

For sample #15061

TVP: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D5191:13.

DVPE: The conversion of the measured Total Vapour Pressure to the corresponding Dry Vapour Pressure Equivalent (DVPE) as described in the ASTM D5191:13 and the U.S. EPA guidelines (40 CFR Part 80, App. E, Method 3), showed one statistical outlier. Both calculated reproducibilities after rejection of the statistical outlier are in agreement with the requirement of ASTM D5191:13.

For sample #15062

RON: This determination was not problematic. Following EN228 (note b) the measured test result for RON should be corrected by subtraction of 0.2 to get the final test result. One laboratory reported that its test result was not corrected. Therefore this result was excluded from the statistical evaluation. No statistical outliers were observed. The calculated reproducibility after rejection of the suspect test result is in agreement with the requirements of ASTM D2699:13b. To try to explain the bimodal distribution observed and to check whether the correction was applied on the ISO5164 results and not on the ASTM D2699 results, a separate evaluation per method was done. The differences were not significant and the possible omission of the correction was obviously not correlated to a specific test method.

MON: This determination was not problematic. Following EN228 (note b) the measured test result for MON should be corrected by subtraction of 0.2 to get the final test result. One laboratory reported that its test result was not corrected. Therefore this result was excluded from the statistical evaluation. No statistical outliers were observed. The calculated reproducibility after rejection of the suspect test result is in agreement with the requirements of ASTM D2700:14. To check whether the correction was applied on the ISO5164 results and not on the ASTM D2699 results, a separate evaluation per method was done. The differences were not significant and the possible omission of the correction was obviously not correlated to a specific test method.

4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the relevant standard and the reproducibility as found for the group of participating laboratories. The assigned values, calculated reproducibilities and reproducibilities, derived from literature standards (in casu ASTM, ISO, EN standards) are compared in the next table.

Parameter	unit	n	average	2.8 * sd	R (lit)
API gravity		18	60.27	0.18	0.61
Aromatics (GC)	%V/V	17	25.48	1.24	1.29
Benzene	%V/V	25	0.53	0.06	0.04
Copper Strip 3 hrs at 50°C	-----	23	1(1A)	n.a.	n.a.
Density at 15°C	kg/m ³	37	737.72	0.60	1.50
Initial Boiling Point	°C	36	31.2	4.0	4.9
10% evaporated	°C	35	46.8	1.7	3.2
50% evaporated	°C	36	67.3	1.6	1.9
90% evaporated	°C	33	149.8	2.4	4.0
Final Boiling Point	°C	37	188.1	5.0	6.8
%Vol at 70°C	%V/V	31	53.5	2.5	1.9
%Vol at 100°C	%V/V	29	63.5	1.4	1.7
%Vol at 150°C	%V/V	29	90.0	1.2	1.2
Doctor test		14	negative	n.a.	n.a.
Existent Gum (solvent washed)	mg/100mL	14	0.6	0.7	2.2
FIA Olefins	%V/V	16	10.9	3.1	3.4
FIA Aromatics	%V/V	17	27.1	5.2	3.7
Lead as Pb	mg/L	14	<2.5	n.a.	n.a.
Manganese as Mn	mg/L	7	<0.25	n.a.	n.a.
Mercaptans as S	%M/M	11	<0.0003	n.a.	n.a.
Olefins by GC	%V/V	12	10.44	0.85	1.74
Oxidation Stability	minutes	12	>360	n.a.	n.a.
Ethanol	%V/V	25	10.2	1.0	0.6
Ethers C5	%V/V	14	<0.2	n.a.	n.a.
Ethers C5 or more C atoms	%V/V	14	<0.2	n.a.	n.a.
Ethers C6 or more C atoms	%V/V	12	<0.2	n.a.	n.a.
Oxygen content	%M/M	23	3.8	0.4	0.4
Sulphur	mg/kg	33	3.4	1.5	1.5

table 4: performance evaluation sample #15060

Parameter	unit	n	average	2.8 * sd	R (lit)
TVP acc.to ASTM D5191	psi	29	12.19	0.21	0.36
DVPE acc.to ASTM D5191	psi	32	11.19	0.26	0.35
DVPE acc.to EPA	psi	25	11.31	0.20	0.35

table 5: performance evaluation sample #15061

Parameter	unit	n	average	2.8 * sd	R (lit)
RON	-----	16	96.30	0.68	0.70
MON	-----	16	85.56	0.93	0.90

table 6: performance evaluation sample #15062

Without further statistical calculations, it can be concluded that for several tests there is a good compliance of the group of participants with the relevant standards. The problematic tests have been discussed in paragraph 4.1.

4.3 COMPARISON OF THE PROFICIENCY TEST OF MAY 2015 WITH PREVIOUS PT

Determination	May 2015	May 2014	May 2013	May 2012	May 2011
Number of reporting labs	41	50	48	40	34
Number of results reported	713	1164	892	831	642
Statistical outliers	20	45	25	30	21
Percentage outliers	2.8%	3.9%	2.8%	3.6%	3.3%

table 7: comparison with previous proficiency tests

In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared against the requirements of the respective standards. The conclusions are given in the following table:

Determination	May 2015	May 2014	May 2013	May 2012	May 2011
API gravity	++	++	+	+	++
Aromatics by GC	+	+	-	+	++
Benzene	-	+/-	-	-	+
Density at 15°C	++	++	+/-	+/-	-
Distillation	++	+	+	+	+
Existent Gum (solvent washed)	++	+/-	+/-	--	n.a
FIA - Olefins	+	+	--	++	--
FIA – Aromatics by FIA	-	-	--	+/-	--
Mercaptans as S	n.e.	++	++	++	++
Olefins by GC	++	++	+	++	++
Oxidation Stability	n.e.	n.e.	n.e.	n.e.	n.e.
Ethanol	-	--	--	++	++
Oxygen content	-	+/-	+/-	+/-	--
Sulphur	+/-	+/-	+/-	-	-
TVP acc.to ASTM D5191	++	+	+	+/-	--
DVPE acc.to ASTM D5191	+	+	+	+/-	-
DVPE acc.to EPA	++	+	+	+/-	-
RON	+/-	-	+	+	+/-
MON	+/-	-	-	-	--

table 8: comparison of the quality of the various determinations against the respective standard requirements



The performance of the determinations against the requirements of the respective standards is listed in the above table. The following performance categories were used:

- ++: group performed much better than the standard
- + : group performed better than the standard
- +/-: group performance equals the standard
- : group performed worse than the standard
- : group performed much worse than the standard
- n.e.: not evaluated

APPENDIX 1

Determination of API gravity on sample #15060;



lab	method	value	mark	z(targ)	remarks
52	D4052	60.2		-0.32	
120	D4052	60.3		0.14	
132	D4052	60.27		0.00	
150	D4052	60.2		-0.32	
158	D4052	60.4		0.60	
159	D4052	60.27		0.00	
169	D4052	60.2		-0.32	
171	D4052	60.3		0.14	
175	D4052	60.3		0.14	
194	D4052	60.3		0.14	
312	D4052	60.31		0.19	
333		----		----	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
340	D4052	60.00	G(0.05)	-1.23	
350		----		----	
381		----		----	
447		----		----	
496	D4052	60.23		-0.18	
511		----		----	
631	D4052	60.21		-0.27	
862	D4052	60.27		0.00	
1033		----		----	
1082		----		----	
1126		----		----	
1161	D4052	60.36		0.42	
1191		----		----	
1203	D4052	60.14		-0.59	
1229		----		----	
1299		----		----	
1459		----		----	
1634	D4052	60.3		0.14	
1706		----		----	
1727		----		----	
1776		----		----	
1810		----		----	
1811		----		----	
2130	D4052	60.286		0.08	
2146		----		----	
normality					
OK					
n					
18					
outliers					
1					
mean (n)					
60.269					
st.dev. (n)					
0.0633					
R(calc.)					
0.177					
R(D4052:11)					
0.611					



Determination of Aromatics by GC on sample #15060; results in %V/V

lab	method	value	mark	z(targ)	remarks
52		----		----	
120	D5769	24.6		-1.92	
132	D5769	24.73		-1.64	
150	D5769	24.9		-1.27	
158	D5769	22.5	G(0.01)	-6.50	
159	D5769	23.89	G(0.05)	-3.47	
169		----		----	
171	D5769	25.3		-0.40	
175		----		----	
194		----		----	
312	ISO22854	26.0		1.13	
333		----		----	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
340		----		----	
350		----		----	
381		----		----	
447		----		----	
496	ISO22854	25.81		0.72	
511		----		----	
631		----		----	
862	ISO22854	25.37		-0.24	
1033		----		----	
1082		----		----	
1126	in house	25.72		0.52	
1161	ISO22854	25.7		0.48	
1191	ISO22854	25.48		0.00	
1203	ISO22854	25.79		0.67	
1229	ISO22854	25.76		0.61	
1299	ISO22854	25.8		0.69	
1459		----		----	
1634		----		----	
1706		----		----	
1727		----		----	
1776	ISO22854	25.76		0.61	
1810		25.6		0.26	
1811	ISO22854	25.97		1.06	
2130	D6730	24.9	C	-1.27	first reported:23.688
2146		----		----	
<u>Only ISO22854 data</u>					
normality		OK		OK	
n		17		11	
outliers		2		0	
mean (n)		25.482		25.731	
st.dev. (n)		0.44365		0.1890	
R(calc.)		1.242		0.529	
R(ISO22854:14)		1.285		1.296	



Compare R(D5769:10) = 2.767



Determination of Benzene on sample #15060; results in %V/V

lab	method	value	mark	z(targ)	remarks
52	INH-14.3	0.54		0.88	
120	D5769	0.53	C	0.18	first reported:0.51
132	D3606	0.570		2.98	
150	D3606	0.51		-1.22	
158	D3606	0.52		-0.52	
159		----		----	
169	D3606	0.5482		1.45	
171	D3606	0.52		-0.52	
175	D3606	0.50		-1.92	
194		----		----	
312	ISO22854	0.54		0.88	
333	EN238	0.49		-2.62	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
340	EN238	0.50		-1.92	
350		----		----	
381		----		----	
447	IP429	0.51218		-1.07	
496	ISO22854	0.545		1.23	
511		----		----	
631		----		----	
862	ISO22854	0.53		0.18	
1033		----		----	
1082		----		----	
1126	in house	0.55		1.58	
1161	ISO22854	0.53		0.18	
1191	ISO22854	0.5		-1.92	
1203	ISO22854	0.52		-0.52	
1229	ISO22854	0.53		0.18	
1299	ISO22854	0.55		1.58	
1459		0.53		0.18	
1634		----		----	
1706		----		----	
1727		----		----	
1776	ISO22854	0.54		0.88	
1810		0.5		-1.92	
1811	ISO22854	0.53		0.18	
2130	D6730	0.551		1.65	
2146		----		----	
normality		OK			
n		25			
outliers		0			
mean (n)		0.527			
st.dev. (n)		0.0202			
R(calc.)		0.056			
R(ISO22854:14)		0.040			

Compare R(D3606) = 0.119





Determination of Copper strip 3hrs/50°C on sample #15060

lab	method	value	mark	z(targ)	remarks
52	D130	1A		----	
120	D130	1A		----	
132	D130	1A		----	
150	D130	1A		----	
158	D130	1A		----	
159	D130	1A		----	
169	D130	1A		----	
171	D130	1A		----	
175	D130	1A		----	
194	D130	1A		----	
312	D130	1A		----	
333				----	
334				----	
335	D130	1		----	
337				----	
338				----	
340	D130	1A		----	
350				----	
381				----	
447	D130	1B		----	
496	D130	1A		----	
511				----	
631	D130	1A		----	
862	D130	1A		----	
1033				----	
1082				----	
1126				----	
1161	D130	1A		----	
1191				----	
1203	ISO2160	1		----	
1229				----	
1299	D130	1A		----	
1459				----	
1634	D130	1A		----	
1706				----	
1727				----	
1776				----	
1810				----	
1811	ISO2160	1		----	
2130	D130	1A		----	
2146				----	
normality					
n		n.a			
outliers		23			
mean (n)		n.a			
st.dev. (n)		1(A)			
R(calc.)		n.a			
R(130:12)		n.a			











Determination of Density at 15°C on sample #15060; results in kg/m³

lab	method	value	mark	z(targ)	remarks
52	D4052	737.8		0.15	
120	D4052	737.5		-0.41	
132	D4052	737.64		-0.15	
150	D4052	737.8		0.15	
158		-----		-----	
159	D4052	737.7		-0.04	
169	D4052	738.2		0.89	
171	D4052	737.6		-0.23	
175	D4052	737.8	C	0.15	first reported: 0.7378 kg/m ³
194	D4052	737.5		-0.41	
312	D4052	737.5		-0.41	
333	ISO12185	737.7		-0.04	
334	ISO12185	737.3		-0.79	
335	ISO12185	737.6		-0.23	
337	ISO12185	738.0		0.52	
338	ISO12185	737.7		-0.04	
340	ISO12185	737.74		0.04	
350	ISO3675	738.5	G(0.05)	1.45	
381		-----		-----	
447	D4052	737.5		-0.41	
496	ISO12185	737.79		0.13	
511		-----		-----	
631	D4052	737.9		0.33	
862	D4052	737.57		-0.28	
1033	IP365	738.1		0.71	
1082	ISO12185	737.7		-0.04	
1126	ISO12185	738.35		1.17	
1161	ISO12185	737.5		-0.41	
1191	ISO12185	737.5		-0.41	
1203	ISO12185	737.66		-0.11	
1229	ISO12185	737.7		-0.04	
1299	D4052	737.8		0.15	
1459	ISO12185	737.65		-0.13	
1634	ISO12185	737.8		0.15	
1706	ISO12185	737.9		0.33	
1727	D4052	737.69		-0.06	
1776	ISO12185	737.6		-0.23	
1810		737.8		0.15	
1811	ISO12185	737.5		-0.41	
2130	D4052	737.6		-0.23	
2146	ISO12185	737.97		0.47	
	normality	suspect			
	n	37			
	outliers	1			
	mean (n)	737.721			
	st.dev. (n)	0.2131			
	R(calc.)	0.597			
	R(ISO12185:96)	1.500			
					Compare R(D4052:11) = 2.322



Determination of Distillation on sample #15060; results in °C

lab	method	mode	IBP	mark	10%eva	mark	50%eva	mark	90%eva	mark	FBP	mark
52	D86	Automated	30.6	46.7		66.9		149.9			186.1	
120	D86	Automated	31.3	46.6		67.4		150.4			187.7	
132	D86	Automated	30.2	46.5		67.2		150.4			188.7	
150	D86	Automated	33.3	47.8		66.8		149.4			189.1	
158	D86	Automated	32.11	46.5		68.17		149.61			189.5	
159	D86	Automated	34.4	48.3		68.1		151.6			187.0	
169	D86	Automated	29.3	46.2		66.7		150.4			187.2	
171	D86	Automated	31.1	47.3		66.3		153.7	R(0.05)		187.9	
175	D86	Automated	31.2	47.6		68.0		149.9			187.9	
194	D86	Automated	31.4	47.7		67.9		149.0			189.0	
312	D86	Automated	29.2	47.4		67.4		150.1			186.2	
333	D86	Automated	29.7	46.6		66.6		149.5			186.2	
334		Automated	----	45.8		66.4		147.7			185.1	
335	D86	Automated	30.5	46.1		66.4		149.5			185.5	
337			----	----		----		----			----	
338	ISO3405	Automated	30.3	46.1		67.0		150.1			188.8	
340	D86	Automated	29.1	47.2		67.2		150.2			184.6	
350	ISO3405	Manual	31.3	46.2		66.3		147.4			187.7	
381			----	----		----		----			----	
447	D86	Automated	29.9	46.6		67.4		149			187.9	
496	D86	Automated	31.6	46.8		66.9		149.5			189.5	
511			----	----		----		----			----	
631	D86	Manual	33.1	44.5	R(0.05)	67.3		150.0			189.0	
862	D86	Automated	32.4	47.3		67.2		150.0			187.9	
1033	IP123	Automated	31.3	47.5		67.1		151.0			187.4	
1082	ISO3405	Automated	28.7	46.4		68.0		150.2			189.1	
1126	D86	Automated	30.6	47.0		68.0		154.1	R(0.05)		191.8	
1161	ISO3405	Automated	32.4	46.4		67.6		152.9	R(0.05)		187.8	
1191	ISO3405	Automated	31.5	46.6		67.2		150.2			186.2	
1203	ISO3405	Automated	33.7	47.4		67.8		151.4			192.4	
1229	ISO3405	Automated	30.1	46.1		67.0		149.7			187.9	
1299	D86	Automated	33.0	----		----		----			186.2	
1459	ISO3405		31.9	46.1		67.1		149.9			187.0	
1634	D86	Automated	29.0	46.3		66.9		150.2			187.0	
1706	D86	Automated	31.7	47.6		68.8		149.9			188.5	
1727			----	----		----		----			----	
1776	ISO3405	Automated	32.2	46.7		66.9		149.8			191.2	
1810		Automated	30.8	47.5		67.8		150.0			189.0	
1811	D86	Automated	30.1	46.4		67.3		149.9			187.7	
2130	D86	Automated	31.8	46.4		67.8		148.3			190.1	
2146	ISO3405	Automated	33.1	46.8		67.1		150.1			190.4	
		OK			OK	OK		not OK		OK		
	normality			n	36	35	36		33		37	
	outliers			0	1	0		3			0	
	mean (n)			31.22	46.81	67.28		149.82			188.06	
	st.dev. (n)			1.416	0.610	0.586		0.858			1.768	
	R(calc.)			3.96	1.71	1.64		2.40			4.95	
	R(D86:12auto)			4.90	3.20	1.88		3.98			6.78	
Compare	R(D86:12manual)			4.44	2.94	2.94		3.70			4.29	









Determination of Distillation on sample #15060; results in %V/V

lab	Method	%evap. at 70°C	mark	%evap. at 100°C	mark	%evap. at 150°C	mark	residue	mark
52	D86	54.1		----		----		1.0	
120	D86	53.3	C	63.3		89.8		1.0	
132	D86	54.1		64.3		89.8		1.0	
150	D86	54.2		63.7		90.3		0.8	
158		----		----		----		1.1	
159		----		----		----		1.0	
169		----		----		----		1.1	
171	D86	52.0		63.6	C	89.8	C	1.5	
175		----		----		----		1.0	
194		----		----		----		0.5	
312	D86	53.6		62.9		90.0		1.0	
333	D86	54.2		63.7		90.2		1.3	
334	D86	53.5		64.0		90.7		0.9	
335	D86	55.2		63.3		90.1		1.0	
337		----		----		----		----	
338	ISO3405	54.7		63.3		90.0		1.0	
340	D86	53.8		63.3		89.9		1.0	
350	ISO3405	53.7		64.2		90.7		0.8	
381		----		----		----		----	
447	D86	53.6		63.5		90.9		1.0	
496	D86	53.9		63.5		90.2		0.8	
511		----		----		----		----	
631	D86	53.0		64.2		89.8		1.1	
862	D86	53.1		63.4		90.0		1.0	
1033	IP123	53.2		60.9	R(0.01)	----		1.2	
1082	ISO3405	54.8		63.3		89.9		0.9	
1126	D86	53.4		63.1		90.0		0.9	
1161	ISO3405	52.3		63.0		89.0		0.8	
1191	ISO3405	54.1		63.7		89.9		1.0	
1203	ISO3405	52.7		62.5		89.5		1.0	
1229	ISO3405	54.4		63.7		90.1		1.1	
1299	D86	52.9		63.3		90.0		1.0	
1459	ISO3405	53.6		63.6		90.0		1.0	
1634	D86	51.3		64.3		89.3		1.0	
1706		----		----		----		1.1	
1727		----		----		----		----	
1776	ISO3405	53.2		63.7		90.1		1.0	
1810		52.0	C	62.6	C	89.4	C	1.0	
1811	D86	54.6		63.6		90.1		1	
2130	D86	52.6		62.7		90.7		1.0	
2146	ISO3405	53.3		62.9		90.0		1.1	
	normality	OK		OK		OK		1.0	
	n	31		29		29			
	outliers	0		1		0			
	mean (n)	53.50		63.46		90.01			
	st.dev. (n)	0.889		0.482		0.413			
	R(calc.)	2.49		1.35		1.16			
	R(D86:12 auto)	1.93		1.73		1.20			
Compare	R(D86:12 man)	4.63		unknown		4.44			

Lab 120 first reported: % evap at 70 °C: 29.6

Lab 171 first reported: % evap at 100 °C: 62.0; % evap at 150 °C: 88.5

Lab 1810 first reported: % evap at 70 °C: no result, % evap at 100 °C: 52.0 %, evap at 150 °C: 62.6





Determination of Doctor test on sample #15060

lab	method	value	mark	z(targ)	remarks
52	D4952	Negative	-----		
120	D4952	Negative	-----		
132	D4952	Negative	-----		
150	D4952	Negative	-----		
158		-----	-----		
159	D4952	Negative	-----		
169		-----	-----		
171	D4952	Negative	-----		
175		-----	-----		
194	D4952	Negative	-----		
312	IP30	Negative	-----		
333	D4952	Negative	-----		
334		-----	-----		
335		-----	-----		
337		-----	-----		
338		-----	-----		
340	D4952	Negative	-----		
350		-----	-----		
381		-----	-----		
447		-----	-----		
496		-----	-----		
511		-----	-----		
631		-----	-----		
862	D4952	Negative	-----		
1033		-----	-----		
1082		-----	-----		
1126		-----	-----		
1161		-----	-----		
1191		-----	-----		
1203	D4952	Negative	-----		
1229		-----	-----		
1299		-----	-----		
1459		-----	-----		
1634		-----	-----		
1706		-----	-----		
1727		-----	-----		
1776		-----	-----		
1810		-----	-----		
1811	D4952	Negative	-----		
2130	D4952	Negative	-----		
2146		-----	-----		
normality		n.a			
n		14			
outliers		n.a			
mean (n)		Negative			
st.dev. (n)		n.a			
R(calc.)		n.a			
R()		n.a			



Determination of Existence Gum (solvent washed) on sample #15060; results in mg/100mL

lab	method	value	mark	z(targ)	remarks
52	D381	0.5		-0.14	
120	D381	<0.5		-----	
132	D381	0.8		0.25	
150	D381	2.4	DG(0.01)	2.33	
158		-----		-----	
159	D381	1.0		0.51	
169		-----		-----	
171	D381	<0.5		-----	
175		-----		-----	
194		-----		-----	
312	D381	0.5		-0.14	
333		-----		-----	
334	D381	0.2		-0.53	
335		-----		-----	
337		-----		-----	
338		-----		-----	
340	D381	<1		-----	
350		-----		-----	
381		-----		-----	
447	D381	2.5	DG(0.01)	2.46	
496	D381	0.4		-0.27	
511		-----		-----	
631	D381	<0.5		-----	
862	D381	0.8		0.25	
1033	IP131	1.90	G(0.01)	1.68	
1082	ISO6246	<1		-----	
1126		-----		-----	
1161	ISO6246	0.74		0.17	
1191	ISO6246	0.3		-0.40	
1203	D381	0.6		-0.01	
1229	ISO6246	<1		-----	
1299	D381	1.0		0.51	
1459		-----		-----	
1634		-----		-----	
1706		-----		-----	
1727		-----		-----	
1776		-----		-----	
1810		0.5		-0.14	
1811	D381	0.6		-0.01	
2130	D381	0.6		-0.01	
2146		-----		-----	
normality		OK			
n		14			
outliers		3			
mean (n)		0.61			
st.dev. (n)		0.238			
R(calc.)		0.67			
R(D381:12)		2.15			





Determination of FIA - Olefins on sample #15060; results in %V/V

lab	method	value	mark	z(targ)	remarks
52		----		----	
120	D1319	11.1		0.14	
132	D1319	10.55		-0.31	
150	D1319	11.6		0.54	
158	D1319	9.8		-0.92	
159	D1319	12.2		1.03	
169	D1319	10.5	C	-0.35	first reported: 25.9
171	D1319	9.1		-1.49	
175	D1319	10.6		-0.27	
194	D1319	12.6		1.36	
312	D1319	12.6		1.36	
333		----		----	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
340	D1319	9.9		-0.84	
350		----		----	
381		----		----	
447	D1319	11.7		0.63	
496	D1319	11.8		0.71	
511		----		----	
631	D1319	9.92		-0.82	
862	D1319	9.51		-1.15	
1033		----		----	
1082		----		----	
1126		----		----	
1161		----		----	
1191		----		----	
1203		----		----	
1229		----		----	
1299		----		----	
1459		----		----	
1634		----		----	
1706		----		----	
1727		----		----	
1776		----		----	
1810		----		----	
1811		----		----	
2130	D1319	11.4		0.38	
2146	D1319	7.0	G(0.05)	-3.20	
normality					
n					
outliers					
mean (n)					
st.dev. (n)					
R(calc.)					
R(D1319:14)					



Determination of FIA – Aromatics on sample #15060; results in %V/V

lab	method	value	mark	z(targ)	remarks
52		----		----	
120	D1319	26.1		-0.78	
132	D1319	26.70		-0.32	
150	D1319	26.1		-0.78	
158	D1319	24.2		-2.21	
159	D1319	29.3		1.65	
169	D1319	25.9	C	-0.93	first reported:10.5
171	D1319	23.9		-2.44	
175	D1319	26.6		-0.40	
194	D1319	28.2		0.81	
312	D1319	28.5		1.04	
333		----		----	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
340	D1319	27.8		0.51	
350		----		----	
381		----		----	
447	D1319	27.6		0.36	
496	D1319	29.0		1.42	
511		----		----	
631	D1319	26.71		-0.31	
862	D1319	26.31		-0.62	
1033		----		----	
1082		----		----	
1126		----		----	
1161		----		----	
1191		----		----	
1203		----		----	
1229		----		----	
1299		----		----	
1459		----		----	
1634		----		----	
1706		----		----	
1727		----		----	
1776		----		----	
1810		----		----	
1811		----		----	
2130	D1319	26.6		-0.40	
2146	D1319	31.6		3.39	
normality					
n					
outliers					
mean (n)					
st.dev. (n)					
R(calc.)					
R(D1319:14)					
OK					
17					
0					
27.12					
1.873					
5.24					
3.70					



Determination of Lead as Pb on sample #15060; results in mg/L

lab	method	value	mark	z(targ)	remarks
52	D3237	<2.5	-----		
120		-----	-----		
132	D3237	<2.5	-----		
150	D3237	<2.5	-----		
158		-----	-----		
159		-----	-----		
169		-----	-----		
171	D3237	<2.5	-----		
175		-----	-----		
194		-----	-----		
312	D3237	<2.5	-----		
333		-----	-----		
334		-----	-----		
335		-----	-----		
337		-----	-----		
338		-----	-----		
340		-----	-----		
350		-----	-----		
381		-----	-----		
447		-----	-----		
496	D3237	<0.1	-----		
511		-----	-----		
631	D3237	<2.5	-----		
862	D3237	<1	-----		
1033		-----	-----		
1082		-----	-----		
1126		-----	-----		
1161		-----	-----		
1191	in house	0.31	-----		
1203	in house	<1	-----		
1229		0	-----		
1299	EN237	<2.5	-----		
1459		<5	-----		
1634		-----	-----		
1706		-----	-----		
1727		-----	-----		
1776		-----	-----		
1810		-----	-----		
1811		-----	-----		
2130	IP352	<2.5	-----		
2146	in house	0.4	-----		
normality					
n					
outliers					
mean (n)					
st.dev. (n)					
R(calc.)					
R(D3237:12)					
Application range: 2.5 - 25 mg/L					

Determination of Manganese as Mn on sample #15060; results in mg/L

lab	method	value	mark	z(targ)	remarks
52	D3831	<0.25	-----	-----	
120		-----	-----	-----	
132	D3831	0.9	-----	-----	false positive test result?
150		-----	-----	-----	
158		-----	-----	-----	
159		-----	-----	-----	
169		-----	-----	-----	
171	D3831	<0.25	-----	-----	
175		-----	-----	-----	
194		-----	-----	-----	
312	D3831	<0.25	-----	-----	
333		-----	-----	-----	
334		-----	-----	-----	
335		-----	-----	-----	
337		-----	-----	-----	
338		-----	-----	-----	
340		-----	-----	-----	
350		-----	-----	-----	
381		-----	-----	-----	
447		-----	-----	-----	
496	EN16136	0.13	-----	-----	
511		-----	-----	-----	
631	D3831	<0.25	-----	-----	
862	D3831	<0.25	-----	-----	
1033		-----	-----	-----	
1082	in house	0.015	-----	-----	
1126		-----	-----	-----	
1161		-----	-----	-----	
1191	EN16136	0.42	-----	-----	
1203	EN16136	<0.5	-----	-----	
1229		-----	-----	-----	
1299		-----	-----	-----	
1459		-----	-----	-----	
1634		-----	-----	-----	
1706		-----	-----	-----	
1727		-----	-----	-----	
1776		-----	-----	-----	
1810		-----	-----	-----	
1811		-----	-----	-----	
2130		-----	-----	-----	
2146	in house	1.8	-----	-----	false positive test result?
	normality	n.a			
	n	7			
	outliers	n.a			
	mean (n)	<0.25			
	st.dev. (n)	n.a			
	R(calc.)	n.a			
	R(D3831:12)	n.a			



Application range: 0.25 – 40 mg/L

Determination of Mercaptans as S on sample #15060; results in %M/M

lab	method	value	mark	z(targ)	remarks
52	D3227	<0.0003	----		
120	D3227	<0.0003	----		
132	D3227	<0.0003	----		
150	D3227	<0.0003	----		
158		----	----		
159	D3227	<0.0001	----		
169		----	----		
171	D3227	<0.0003	----		
175		----	----		
194		----	----		
312	D3227	<0.0003	----		
333		----	----		
334	D3227	0.00005	----		
335		----	----		
337		----	----		
338		----	----		
340	D3227	0.0003	----		
350		----	----		
381		----	----		
447		----	----		
496	D3227	0.00005	----		
511		----	----		
631		----	----		
862		----	----		
1033		----	----		
1082		----	----		
1126		----	----		
1161		----	----		
1191		----	----		
1203	UOP163	0.000060	----		
1229		----	----		
1299		----	----		
1459		----	----		
1634		----	----		
1706		----	----		
1727		----	----		
1776		----	----		
1810		----	----		
1811		----	----		
2130	D3227	0.0001	----		
2146		----	----		
normality		n.a			
n		11			
outliers		n.a			
mean (n)		<0.0003			
st.dev. (n)		n.a			
R(calc.)		n.a			
R(D3227:13)		n.a			
Application range: 0.0003 – 0.01% M/M					

Determination of Olefins by GC on sample #15060; results in %V/V

lab	method	value	mark	z(targ)	remarks
52		----		----	
120		----		----	
132		----		----	
150		----		----	
158		----		----	
159		----		----	
169		----		----	
171		----		----	
175		----		----	
194		----		----	
312	ISO22854	10.3		-0.23	
333		----		----	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
340		----		----	
350		----		----	
381		----		----	
447		----		----	
496	ISO22854	10.21		-0.37	
511		----		----	
631		----		----	
862	ISO22854	10.99		0.89	
1033		----		----	
1082		----		----	
1126	in house	10.44		0.00	
1161	ISO22854	10.23		-0.34	
1191	ISO22854	10.82		0.61	
1203	ISO22854	10.60		0.26	
1229	ISO22854	10.04		-0.64	
1299	ISO22854	10.5		0.10	
1459		----		----	
1634		----		----	
1706		----		----	
1727		----		----	
1776	ISO22854	9.99		-0.72	
1810		10.5		0.10	
1811	ISO22854	10.66		0.35	
2130	D6730	8.4	C,G(0.01)	-3.28	first reported:8.288
2146		----		----	
normality		OK			
n		12			
outliers		1			
mean (n)		10.440			
st.dev. (n)		0.3026			
R(calc.)		0.847			
R(ISO22854:14)		1.740			





Determination of Oxidation Stability on sample #15060; results in minutes

lab	method	value	mark	z(targ)	remarks
52	D525	>900	-----		
120		-----	-----		
132	D525	>954	-----		
150	D525	>900	-----		
158		-----	-----		
159		-----	-----		
169		-----	-----		
171	D525	241	-----	False positive test result?	
175		-----	-----		
194		-----	-----		
312	D525	>900	-----		
333		-----	-----		
334		-----	-----		
335		-----	-----		
337		-----	-----		
338		-----	-----		
340	ISO7536	>900	-----		
350		-----	-----		
381		-----	-----		
447	D525	>900	-----		
496	D525	>900	-----		
511		-----	-----		
631		-----	-----		
862	D525	>900	-----		
1033		-----	-----		
1082		-----	-----		
1126		-----	-----		
1161	ISO7536	>900	-----		
1191		-----	-----		
1203	ISO7536	>900	-----		
1229		-----	-----		
1299	D525	>960	-----		
1459		-----	-----		
1634		-----	-----		
1706		-----	-----		
1727		-----	-----		
1776		-----	-----		
1810		-----	-----		
1811		-----	-----		
2130	D525	>900	-----		
2146		-----	-----		
normality		n.a			
n		12			
outliers		n.a			
mean (n)		>360			
st.dev. (n)		n.a			
R(calc.)		n.a			
R(D525:12a)		n.a			

Determination of Ethanol on sample #15060; results in %V/V

lab	method	value	mark	z(targ)	remarks
52		----		----	
120	D5599	9.76		-2.14	
132	D5599	10.34		0.53	
150	D5599	10.69		2.14	
158	D5599	10.36		0.62	
159		----		----	
169	D4815	10.82		2.74	
171	D4815	10.38	C	0.72	first reported: 11.21
175	D5599	10.23		0.03	
194		----		----	
312	ISO22854	10.10		-0.57	
333	EN13132	9.6		-2.88	
334	EN1601	9.2	G(0.05)	-4.72	
335	EN1601	9.6		-2.88	
337	EN13132	9.7		-2.41	
338		----		----	
340		----		----	
350	EN13132	10.49		1.22	
381		----		----	
447		----		----	
496	ISO22854	10.135		-0.41	
511		----		----	
631		----		----	
862	D4815	10.419		0.90	
1033		----		----	
1082		----		----	
1126	in house	8.96	G(0.05)	-5.82	
1161	EN22854	10.1		-0.57	
1191	EN1601	11.001		3.58	
1203	ISO22854	10.17		-0.25	
1229	EN1601	10.3275		0.47	
1299	ISO22854	10.61		1.78	
1459		10.35		0.58	
1634		----		----	
1706	EN13132	9.99		-1.08	
1727		----		----	
1776	ISO22854	10.34		0.53	
1810		9.9		-1.49	
1811	EN22854	10.21		-0.07	
2130	D6730	9.989		-1.08	
2146		----		----	
normality		OK			
n		25			
outliers		2			
mean (n)		10.22			
st.dev. (n)		0.357			
R(calc.)		1.00			
R(ISO22854:14)		0.61			

Compare R(D5599:10) = 1.73, R(EN1601:14)=R(EN13132:00)=0.8



Determination of Ethers (C5, C5 or more C atoms and C6 and more) on sample #15060; results in %V/V

lab	Method	C5	mark	C5 or more	mark	C6 and more	mark
52		----		----		----	
120	D5599	<0.01		<0.01		<0.01	
132	D5599	<0.10		<0.10		<0.10	
150	D5599	<0.10		<0.10		<0.10	
158	D5599	<0.01		<0.01		<0.01	
159		----		----		----	
169		----		----		----	
171	D5599	<0.10		<0.10		<0.10	
175		----		----		----	
194		----		----		----	
312	ISO22854	0.05		0.1		<0.1	
333		----		----		----	
334	EN1601	0		0		0	
335		----		----		----	
337	EN13132	<0.17		<0.17		<0.17	
338		----		----		----	
340		----		----		----	
350		----		0.09		----	
381		----		----		----	
447		----		----		----	
496	ISO22854	0.090		0.140		0.050	
511		----		----		----	
631		----		----		----	
862	D4815	<0.2		----		----	
1033		----		----		----	
1082		----		----		----	
1126		----		----		----	
1161		----		----		----	
1191		----		----		----	
1203	ISO22854	0.05		0.09		0.04	
1229		----		----		----	
1299		----		<0.8		----	
1459		----		----		----	
1634		----		----		----	
1706		----		----		----	
1727		----		----		----	
1776	ISO22854	<0.20		<0.20		<0.20	
1810		----		----		----	
1811		----		----		----	
2130	D6730	<0.1		<0.1		<0.1	
2146		----		----		----	
normality		n.a		n.a		n.a	
n		13		13		12	
outliers		n.a		n.a		n.a	
mean (n)		<0.2		<0.2		<0.2	
st.dev. (n)		n.a		n.a		n.a	
R(calc.)		n.a		n.a		n.a	
R(ISO22854:14)		n.a		n.a		n.a	

Determination of Oxygenates on sample #15060; results in %V/V

lab	Method	DIPE	mark	ETBE	mark	i-BuOH	mark	IPA	mark	MeOH	mark	MTBE	mark
52		----		----		----		----		----		----	
120	D5599	<0.01		<0.01		<0.01		<0.01		0.053		<0.01	
132	D5599	<0.10		<0.10		<0.10		<0.10		<0.10		<0.10	
150	D5599	<0.10		<0.10		<0.10		<0.10		<0.10		<0.10	
158	D5599	<0.01		<0.01		<0.01		<0.01		<0.01		<0.01	
159		----		----		----		----		----		----	
169		----		----		----		----		----		----	
171	D4815	<0.20		<0.20		<0.10		<0.10		<0.10		<0.10	
175		----		----		----		----		----		0.22	
194		----		----		----		----		----		----	
312	ISO22854	<0.01		0.04		<0.01		<0.01		<0.01		0.04	
333		----		<0.17		<0.17		<0.17		<0.17		<0.17	
334	EN1601	0		0		0		0		0		0	
335		----		----		----		----		----		----	
337	EN13132	<0.17		0.05	C	<0.17		<0.17		<0.17		<0.17	
338		----		----		----		----		----		----	
340		----		----		----		----		----		----	
350		----		0.04		----		----		----		0.05	
381		----		----		----		----		----		----	
447		----		----		----		----		----		----	
496	ISO22854	<0.010		0.050		<0.10		<0.10		0.065		0.090	
511		----		----		----		----		----		----	
631		----		----		----		----		----		----	
862	D4815	<0.2		<0.2		<0.2		<0.2		<0.2		<0.2	
1033		----		----		----		----		----		----	
1082		----		----		----		----		----		----	
1126		----		0.04		----		----		----		0.05	
1161		----		----		----		----		----		0.1	
1191		----		0.035		----		----		0.048		0.044	
1203	ISO22854	<0.02		0.04		<0.02		<0.02		<0.02		0.05	
1229	EN1601	0		0.036		0		0		0		0.043	
1299		----		<0.8		<0.8		<0.8		<0.8		<0.8	
1459		----		0.04		----		----		----		----	
1634		----		----		----		----		----		----	
1706		----		----		----		----		----		----	
1727		----		----		----		----		----		----	
1776	ISO22854	<0.20		0.04		<0.20		<0.20		<0.20		0.05	
1810		----		----		----		----		----		0.05	
1811		----		0.04		----		----		0.0		0.05	
2130	D6730	<0.1		<0.1		<0.1		<0.1		0.089		<0.1	
2146		----		----		----		----		----		----	
normality													
n		n.a		n.a		n.a		n.a		n.a		n.a	
		14		20		15		15		17		21	
outliers		n.a		n.a		n.a		n.a		n.a		n.a	
mean (n)		<0.2		<0.2		<0.2		<0.2		<0.2		<0.2	
st.dev. (n)		n.a		n.a		n.a		n.a		n.a		n.a	
R(calc.)		n.a		n.a		n.a		n.a		n.a		n.a	
R(ISO22854:14)		n.a		n.a		n.a		n.a		n.a		n.a	

Lab 33 first reported ETBE: <0.17



Determination of Oxygenates on sample #15060; results in %V/V

Lab	Method	TAME	mark	t-BuOH	mark	Other oxygenates	mark
52		----		----		----	
120	D5599	<0.01		<0.01		<0.01	
132	D5599	<0.10		<0.10		<0.10	
150	D5599	<0.10		<0.10		<0.10	
158	D5599	<0.01		<0.01		<0.01	
159		----		----		----	
169		----		----		----	
171	D4815	<0.10		<0.10		<0.10	
175		----		----		----	
194		----		----		----	
312	ISO22854	<0.01		<0.01		<0.1	
333		----		<0.17		----	
334	EN1601	0		0		0	
335		----		----		----	
337	EN13132	<0.17		<0.17		<0.17	
338		----		----		----	
340		----		----		----	
350		----		----		----	
381		----		----		----	
447		----		----		----	
496	ISO22854	<0.010		<0.1		<0.10	
511		----		----		----	
631		----		----		----	
862	D4815	<0.2		<0.2		----	
1033		----		----		----	
1082		----		----		----	
1126		----		----		----	
1161		----		----		----	
1191		<0.05		----		----	
1203	ISO22854	<0.02		<0.02		0.03	
1229	EN1601	0		0		0	
1299		----		<0.8		<0.8	
1459		----		----		----	
1634		----		----		----	
1706		----		----		----	
1727		----		----		----	
1776	ISO22854	<0.20		<0.20		<0.20	
1810		----		----		----	
1811		----		----		----	
2130	D6730	<0.1		<0.1		<0.1	
2146		----		----		----	
normality		n.a		n.a		n.a	
n		15		15		13	
outliers		n.a		n.a		n.a	
mean (n)		<0.2		<0.2		<0.2	
st.dev. (n)		n.a		n.a		n.a	
R(calc.)		n.a		n.a		n.a	
R(ISO22854:14)		n.a		n.a		n.a	

Determination of Oxygen content on sample #15060; results in %M/M



lab	method	value	mark	z(targ)	remarks
52		----		----	
120	D5599	3.68		-1.18	
132	D5599	3.9		0.49	
150	D5599	3.99		1.17	
158	D5599	3.87		0.26	
159		----		----	
169		----		----	
171	D5599	4.0		1.24	
175	D5599	3.88		0.33	
194		----		----	
312	ISO22854	3.80		-0.27	
333	EN13132	3.61		-1.71	
334	EN1601	3.43		-3.07	
335		----		----	
337	EN13132	3.66		-1.33	
338		----		----	
340		----		----	
350	EN13132	3.92		0.64	
381		----		----	
447		----		----	
496	ISO22854	3.855		0.15	
511		----		----	
631		----		----	
862	D4815	3.891		0.42	
1033		----		----	
1082		----		----	
1126	in house	9.05	R(0.01), E	39.46	calculation error was observed, iis calculated 3.37%
1161	ISO22854	3.8		-0.27	
1191	EN1601	4.158		2.44	
1203	ISO22854	3.83		-0.04	
1229	EN1601	3.89		0.41	
1299	ISO22854	3.99		1.17	
1459		3.88		0.33	
1634		----		----	
1706		----		----	
1727		----		----	
1776	ISO22854	3.87		0.26	
1810		3.7		-1.03	
1811	ISO22854	3.84		0.03	
2130	D6730	3.780		-0.42	
2146		----		----	
normality					
n		suspect			
outliers		23			
mean (n)		1.			
st.dev. (n)		3.84			
R(calc.)		0.150			
R(EN228:13(k))		0.42			
		0.37			

Application range ISO22854:14 <3%M/M (or <3.7%M/M, see EN228 (f))





Determination of Sulphur on sample #15060; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52	D5453	3.0		-0.75	
120	D2622	3.8		0.74	
132	D2622	2.94		-0.87	
150	D2622	3.9		0.93	
158	D5453	2.7		-1.31	
159	D5453	4.0		1.11	
169	D5453	3.3		-0.19	
171	D5453	3.5		0.18	
175	D5453	3.3		-0.19	
194	D5453	3.1		-0.57	
312	ISO20846	3.2		-0.38	
333	ISO20846	3.4		-0.01	
334	ISO20846	3.2		-0.38	
335	ISO20846	3.52		0.22	
337	ISO20846	3.5		0.18	
338	ISO20846	2.9		-0.94	
340	ISO20846	2.6		-1.50	
350		----		----	
381		----		----	
447		----		----	
496	ISO20846	2.58		-1.54	
511		----		----	
631		----		----	
862	D5453	3.4		-0.01	
1033		----		----	
1082		----		----	
1126	IOS20846	4.14		1.37	
1161	ISO20846	3.15		-0.47	
1191	ISO20846	3.3		-0.19	
1203	ISO20846	2.9	C	-0.94	first reported:1.5
1229	ISO20846	3.7		0.55	
1299	ISO20884	4.3		1.67	
1459		4.7		2.42	
1634	ISO20846	3.6		0.37	
1706	ISO20884	3.9		0.93	
1727		----		----	
1776	ISO20846	2.7		-1.31	
1810		3.1		-0.57	
1811	ISO20846	3.35		-0.10	
2130	IP490	4.45		1.95	
2146	ISO20846	3.2		-0.38	
normality					
n					
outliers					
mean (n)					
st.dev. (n)					
R(calc.)					
R(ISO20846:11)					





Determination of Total Vapour Pressure on sample #15061; results in psi

lab	method	value	mark	z(targ)	remarks
52	D5191	12.23		0.35	
53	D5191	12.33		1.13	
120	D5191	12.18		-0.04	
132	D5191	12.27		0.66	
150	D5191	12.15		-0.28	
158	D5191	12.28		0.74	
159	D5191	12.17		-0.12	
169	D5191	12.26		0.58	
171	D5191	12.20		0.11	
175	D5191	12.26		0.58	
194	D5191	12.07		-0.90	
312	D5191	12.19		0.04	
333	D5191	12.15		-0.28	
334		-----		-----	
335	D5191	12.20		0.11	
337	D5191	12.141		-0.35	
338	D5191	12.17		-0.12	
340		-----		-----	
350	EN13016-1	12.285	C	0.78	first reported: 11.458
357	D5191	12.27		0.66	
381		-----		-----	
447		-----		-----	
496	D5191	12.24		0.43	first reported: 11.661
631	D5191	12.081	C	-0.81	
862	D5191	12.065		-0.94	
1033		-----		-----	
1082	EN13016-1	12.04	C	-1.13	first reported: 80.9 kPa
1161	EN13016-1	12.08		-0.82	
1191	EN13016-1	12.10		-0.67	
1203	EN13016-1	11.92	R(0.05)	-2.07	
1229	EN13016-1	12.17		-0.12	
1299		-----		-----	
1459		-----		-----	
1634	EN13016-1	12.17		-0.12	
1776	EN13016-1	12.2		0.11	
1810	D5191	11.92	R(0.05)	-2.07	
2130	D5191	12.186		0.01	
2146	EN13016-1	12.237		0.40	
normality					
n					
outliers					
mean (n)					
st.dev. (n)					
R(calc.)					
R(D5191:13)					





Determination of DVPE (ASTM D5191 calculation) on sample #15061; results in psi

lab	method	value	mark	z(targ)	remarks
52	D5191	11.25		0.51	
53	D5191	11.34		1.24	
120	D5191	11.20		0.11	
132	D5191	11.29		0.84	
150	D5191	11.18		-0.05	
158		----		----	
159	D5191	11.196		0.08	
169	D5191	11.28		0.76	
171	D5191	11.23		0.35	
175	D5191	11.28		0.76	
194		----		----	
312	D5191	11.22		0.27	
333	D5191	11.18		-0.05	
334	D5191	11.04		-1.17	
335	D5191	11.21		0.19	
337	D5191	11.168		-0.14	
338	D5191	11.20		0.11	
340	EN13016-1	11.12		-0.53	
350	EN13016-1	11.307	C	0.97	first reported: 10.509
357	D5191	11.29		0.84	
381		----		----	
447	D5191	11.23		0.35	
496	D5191	11.264		0.63	
631	D5191	11.105	C	-0.65	first reported: 10.705
862	D5191	11.090		-0.77	
1033		----		----	
1082	EN13016-1	11.07		-0.93	
1161	EN13016-1	11.11		-0.61	
1191	EN13016-1	11.12		-0.53	
1203		10.96		-1.81	
1229		----		----	
1299	D5191	11.18		-0.05	
1459		----		----	
1634	EN13016-1	11.20		0.11	
1776	EN13016-1	11.2		0.11	
1810		10.96		-1.81	
2130	D5191	11.22		0.27	
2146	EN13016-1	11.261		0.60	
 normality					
OK					
n					
32					
outliers					
0					
mean (n)					
11.186					
st.dev. (n)					
0.0931					
R(calc.)					
0.261					
R(D5191:13)					
0.349					





Determination of DVPE (EPA calculation) on sample #15061; results in psi

lab	method	value	mark	z(targ)	remarks
52	D5191	11.35		0.33	
53		----		----	
120	D5191	11.30		-0.07	
132	D5191	11.39		0.65	
150	D5191	11.26		-0.39	
158	D5191	11.39		0.65	
159	D5191	11.287		-0.18	
169	D5191	11.37		0.49	
171	D5191	11.31		0.01	
175	D5191	11.37		0.49	
194	D5191	11.18		-1.03	
312	D5191	11.31		0.01	
333	D5191	11.28		-0.23	
334		----		----	
335	D5191	11.31		0.01	
337	D5191	11.260		-0.39	
338	D5191	11.29		-0.15	
340		----		----	
350	EN13016-1	11.397	C	0.70	first reported: 10.607
357	D5191	11.38		0.57	
381		----		----	
447		----		----	
496	D5191	11.354		0.36	
631	D5191	11.201	C	-0.87	first reported: 10.801
862	D5191	11.182		-1.02	
1033		----		----	
1082		----		----	
1161		----		----	
1191		----		----	
1203		11.05	G(0.05)	-2.07	
1229	EN13016-1	11.20		-0.87	
1299		----		----	
1459		----		----	
1634	EN13016-1	11.289		-0.16	
1776	EN13016-1	11.3		-0.07	
1810		----		----	
2130	D5191	11.31		0.01	
2146	EN13016-1	11.462		1.22	
normality					
n					
25					
outliers					
1					
mean (n)					
11.309					
st.dev. (n)					
0.0718					
R(calc.)					
0.201					
R(D5191:13)					
0.350					





Determination of RON on sample #15062

lab	method	value	mark	z(targ)	remarks
52		----		----	
120	D2699	96.5		0.82	
132	D2699	96.5		0.82	
159	D2699	96.6		1.22	
171	D2699	96.3		0.02	
175	D2699	96.1		-0.78	
312	D2699	96.4		0.42	
334	D2699	96.1		-0.78	
340	D2699	96.0		-1.18	
381		----		----	
447	D2699	96.3		0.02	
631	D2699	96.19		-0.42	
1082	ISO5164	96.5		0.82	
1161	ISO5164	96.1		-0.78	
1191	ISO5164	96.0		-1.18	
1229	ISO5164	96.333		0.15	
1299	D2699	96.0		-1.18	
1776	ISO5164	96.4	ex	0.42	did not subtract the correction of 0.2 for calculation of the final result
2130	D2699	96.8		2.02	
<u>Only D2699 data</u>					
normality	OK			OK	<u>Only ISO5164 data</u>
n	16			12	not OK
outliers	0 + 1 excl.			0	4
mean (n)	96.30			96.32	1
st.dev. (n)	0.241			0.252	96.23
R(calc.)	0.68			0.71	0.226
R(D2699:13b)	0.70			0.70	0.63
					0.70



Determination of MON on sample #15062

lab	method	value	mark	z(targ)	remarks
52	D2700	85.8		0.73	
120	D2700	85.6		0.11	
132	D2700	85.6		0.11	
159	D2700	86.3		2.29	
171	D2700	85.4		-0.51	
175	D2700	85.1		-1.45	
312	D2700	85.4		-0.51	
334	D2700	85.8		0.73	
340		-----		-----	
381		-----		-----	
447	D2700	85.3		-0.82	
631	D2700	85.30		-0.82	
1082	ISO5163	86.1		1.67	
1161	ISO5163	85.6		0.11	
1191	ISO5163	85.3		-0.82	
1229	ISO5163	85.833		0.84	
1299	D2700	85.2		-1.13	
1776	ISO5163	86.5	ex	2.91	did not subtract the correction of 0.2 for calculation of the final result
2130	D2700	85.4		-0.51	
<hr/>					
normality		OK		Only D2700 data	Only ISO5163 data
n		16		not OK	not OK
outliers		0 + 1 excl.		12	4
mean (n)		85.56		0	1
st.dev. (n)		0.332		85.52	85.71
R(calc.)		0.93		0.330	0.340
R(D2700:14)		0.90		0.92	0.95
				0.90	0.90



APPENDIX 2:**Z-scores of Distillation**

lab	IBP	10%eva	50%eva	90%eva	FBP	%volat70°C	%volat100°C	%volat150°C
52	-0.35	-0.10	-0.56	0.05	-0.81	0.88	----	----
120	0.05	-0.19	0.18	0.40	-0.15	-0.29	-0.25	-0.48
132	-0.58	-0.27	-0.11	0.40	0.26	0.88	1.37	-0.48
150	1.19	0.86	-0.71	-0.30	0.43	1.02	0.40	0.68
158	0.51	-0.27	1.33	-0.15	0.59	----	----	----
159	1.82	1.30	1.23	1.25	-0.44	----	----	----
169	-1.10	-0.54	-0.86	0.40	-0.35	----	----	----
171	-0.07	0.43	-1.46	2.72	-0.07	-2.17	0.23	-0.48
175	-0.01	0.69	1.08	0.05	-0.07	----	----	----
194	0.10	0.78	0.93	-0.58	0.39	----	----	----
312	-1.15	0.51	0.18	0.19	-0.77	0.15	-0.90	-0.02
333	-0.87	-0.19	-1.01	-0.23	-0.77	1.02	0.40	0.45
334	----	-0.89	-1.31	-1.49	-1.22	0.00	0.88	1.62
335	-0.41	-0.62	-1.31	-0.23	-1.06	2.47	-0.25	0.22
337	----	----	----	----	----	----	----	----
338	-0.53	-0.62	-0.41	0.19	0.31	1.75	-0.25	-0.02
340	-1.21	0.34	-0.11	0.26	-1.43	0.44	-0.25	-0.25
350	0.05	-0.54	-1.46	-1.70	-0.15	0.29	1.20	1.62
381	----	----	----	----	----	----	----	----
447	-0.75	-0.19	0.18	-0.58	-0.07	0.15	0.07	2.08
496	0.22	-0.01	-0.56	-0.23	0.59	0.58	0.07	0.45
511	----	----	----	----	----	----	----	----
631	1.07	-2.02	0.03	0.12	0.39	-0.72	1.20	-0.48
862	0.67	0.43	-0.11	0.12	-0.07	-0.58	-0.09	-0.02
1033	0.05	0.60	-0.26	0.83	-0.27	-0.43	-4.13	----
1082	-1.44	-0.36	1.08	0.26	0.43	1.89	-0.25	-0.25
1126	-0.35	0.16	1.08	3.01	1.54	-0.14	-0.57	-0.02
1161	0.67	-0.36	0.48	2.16	-0.11	-1.74	-0.74	-2.35
1191	0.16	-0.19	-0.11	0.26	-0.77	0.88	0.40	-0.25
1203	1.42	0.51	0.78	1.11	1.79	-1.16	-1.55	-1.18
1229	-0.64	-0.62	-0.41	-0.09	-0.07	1.31	0.40	0.22
1299	1.02	----	----	----	-0.77	-0.87	-0.25	-0.02
1459	0.39	-0.62	-0.26	0.05	-0.44	0.15	0.23	-0.02
1634	-1.27	-0.45	-0.56	0.26	-0.44	-3.19	1.37	-1.65
1706	0.27	0.69	2.27	0.05	0.18	----	----	----
1727	----	----	----	----	----	----	----	----
1776	0.56	-0.10	-0.56	-0.02	1.30	-0.43	0.40	0.22
1810	-0.24	0.60	0.78	0.12	0.39	-2.17	-1.38	-1.42
1811	-0.64	-0.36	0.03	0.05	-0.15	1.60	0.23	0.22
2130	0.33	-0.36	0.78	-1.07	0.84	-1.30	-1.22	1.62
2146	1.07	-0.01	-0.26	0.19	0.97	-0.29	-0.90	-0.02

APPENDIX 3**Number of participants per country**

1 lab in AUSTRIA
2 labs in CANADA
1 lab in CHINA, People's Republic
2 labs in CZECH REPUBLIC
5 labs in FINLAND
7 labs in FRANCE
1 lab in GERMANY
1 lab in HUNGARY
2 labs in NETHERLANDS
1 lab in PERU
1 lab in PHILIPPINES
1 lab in PORTUGAL
1 lab in SERBIA
3 labs in SPAIN
1 lab in SWEDEN
1 lab in TURKEY
3 labs in UNITED KINGDOM
9 labs in UNITED STATES OF AMERICA

APPENDIX 4**Abbreviations:**

C	= final result after checking of first reported suspect result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
R(0.01)	= outlier in Rosner outlier test
R(0.05)	= straggler in Rosner outlier test
ex	= excluded from calculations
E	= error in calculations
n.a.	= not applicable
n.e.	= not evaluated
W	= withdrawn
fr.	= first reported
U	= reported in different unit
SDS	= Safety Data Sheet

Literature:

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, April 2014
- 2 ASTM E178-02
- 3 ASTM E1301-03
- 4 ISO 5725-86
- 5 ISO 5725, parts 1-6, 1994
- 6 ISO13528-05
- 7 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 8 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 9 IP 367/84
- 10 DIN 38402 T41/42
- 11 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 12 J.N. Miller, Analyst, 118, 455, (1993)
- 13 Analytical Methods Committee Technical Brief, No4 January 2001
- 14 The Royal Society of Chemistry 2002, Analyst 2002, 127 page1359-1364, P.J. Lowthian and M. Thompson. (see <http://www.rsc.org/suppdata/an/b2/b205600n/>)
- 15 H. Verplaetse and M. Lacourt, Accred Qual Assur (2006) 11:521-522
- 16 Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier Procedure, *Technometrics*, 25(2), pp. 165-172, (1983)